

PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:

Haven Metrology Service 13720 172nd Avenue, Grand Haven, MI 49417

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025:2005

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated January 2009):

Dimensional Inspection (As detailed in the supplement)

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen

Initial Accreditation Date:

Issue Date:

Expiration Date:

June 27, 2010

August 30, 2016

September 30, 2018

Accreditation No.:

Certificate No.:

67643

L16-357

President/Operations Manager
Perry Johnson Laboratory

Perry Johnson Laboratory Accreditation, Inc. (PJLA) 755 W. Big Beaver, Suite 1325 Troy, Michigan 48084

The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: www.pjlabs.com





Certificate of Accreditation: Supplement

Haven Metrology Service 13720 172nd Avenue, Grand Haven, MI 49417 Contact Name: Jack Feddema Phone: 616-607-8095

Accreditation is granted to the facility to perform the following testing:

| FIELD OF TEST | ITEMS, MATERIALS OR PRODUCTS TESTED | SPECIFIC TESTS OR PROPERTIES | SPECIFICATION, STANDARD METHOD | RANGE (WHERE APPROPRIATE) AND |
|------------------|-------------------------------------|---------------------------------|-----------------------------------|----------------------------------|
| OF IESI | FRODUCTS TESTED | MEASURED | OR TECHNIQUE USED | DETECTION LIMIT |
| Dimensional | 2 Dimensional and | 2 Dimensional and 3 | Customer Supplied | CMM |
| Inspection | 3 Dimensional | Dimensional | Dimensional | 36 in x 42 in x 80 in |
| | Manufactured Products | Features for Size, | Information | DL = 0.000 1 in |
| | and Components | Location, and | ANSI Y14.5-M | |
| | | Orientation | | |
| | | 2 Dimensional | | Vision Measuring |
| | | Features for Size, | | System |
| | | Location, and | | 18 in x 24 in |
| | | Orientation | | DL = 0.000 1 in |
| | | 2 Dimensional | | Micrometers |
| | | Features for Size | | 0 in to 6 in |
| | | using Micrometers | | DL = 0.000 1 in |
| | | 2 Dimensional | | Gage Pins |
| | | Features for Size | | 0.25 in to 1.0 in |
| | | | | DL = 0.001 in |

