



# PERRY JOHNSON LABORATORY ACCREDITATION, INC.

## Certificate of Accreditation

*Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:*

***Haven Metrology Service***  
13694 172<sup>nd</sup> Avenue, Grand Haven, MI 49417

*(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:*

**ISO/IEC 17025:2017**

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

***Dimensional Calibration***  
*(As detailed in the supplement)*

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen  
President

*Initial Accreditation Date:*

June 27, 2010

*Issue Date:*

August 03, 2022

*Expiration Date:*

October 31, 2024

*Accreditation No.*

67643

*Certificate No.:*

L22-528

Perry Johnson Laboratory  
Accreditation, Inc. (PJLA)  
755 W. Big Beaver, Suite 1325  
Troy, Michigan 48084

*The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: [www.pjllabs.com](http://www.pjllabs.com)*



# Certificate of Accreditation: Supplement

## Haven Metrology Service

13694 172<sup>nd</sup> Avenue, Grand Haven, MI 49417  
Contact Name: Kyle Brummans Phone: 616-607-8095

Accreditation is granted to the facility to perform the following calibrations:

### Dimensional

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Customer Supplied Fixtures & Gages <sup>F</sup>	X = 700 mm Y = 1 000 mm Z = 700 mm	(16.59 + 0.039L) $\mu$ m	Contact CMM Zeiss Brown & Sharp Sheffield
	X= 1 500 mm Y= 2 800 mm Z= 1 400 mm	(38.38 + 0.043L) $\mu$ m	Global GPG 41
	X= 1 500 mm Y= 600 mm Z= 300 mm	(9.01+ 0.027L) $\mu$ m	Optical CMM: Hexagon Optiv GPG 41
Customer Supplied Fixture and Gages <sup>FO</sup>	Radius 1 800 mm	(40.99+ 0.045L) $\mu$ m	Articulated Arm: Faro Quantum S GPG 41
	Volume 75 mm <sup>3</sup> Volume 600 mm <sup>3</sup>	(6.79 + 0.013L) $\mu$ m (21.66 +.027L) $\mu$ m	Non-Contact Structured Light Systems: Zeiss Comet L3D GPG 41
	Up to 0.16 m	(29.19 + 0.501L) $\mu$ m	Micrometers GPG 40

1. The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represents the smallest measurement uncertainty attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is typically expressed at a confidence level of 95 % using a coverage factor *k* (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.
2. The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.
3. The presence of a superscript F means that the laboratory performs calibration of the indicated parameter at its fixed location. Example: Outside Micrometer<sup>F</sup> would mean that the laboratory performs this calibration at its fixed location.
4. The presence of a superscript FO means that the laboratory performs calibration of the indicated parameter both at its fixed location and onsite at customer locations. Example: Outside Micrometer<sup>FO</sup> would mean that the laboratory performs this calibration at its fixed location and onsite at customer locations.
5. The term L represents length inches or millimeters as appropriate to the uncertainty statement.